



## TERMINAL CLEANING FOR EBOLA VIRUS DISEASE CONTAMINATED PATIENT-CARE AREAS

TECHNICAL INFORMATION PAPER No. 13-033-0616

### PURPOSE.

To provide guidance for terminal disinfection and cleaning of patient-care areas used to treat Ebola Virus Disease (EVD) patients. This document does not change any existing Department of Defense (DoD/DOD) directives, policies, or procedures related to normal waste management in DOD medical treatment facilities (MTFs).

### APPLICABILITY.

This technical information paper (TIP) addresses room and area terminal cleaning procedures conducted in EVD patient treatment rooms after patient discharge in order to protect personnel and prepare the room for future patient care. The TIP distinguishes between rooms used during initial entry into the MTF by patients under investigation for Ebola (such as emergency rooms, clinics, or transport vehicles, and isolation rooms used for prolonged EVD patient treatment). This information is applicable to garrison MTFs, clinics, and off-post clinics. Federal, state, or local public health authorities may establish additional requirements. This document is not a step-by-step instruction and requires personnel to read its contents entirely prior to beginning any onsite disinfection activities.

### REFERENCES.

See Appendix A for complete reference information.

### BACKGROUND.

The threat of EVD infection warrants stringent terminal cleaning procedures after an EVD patient is discharged from an MTF. Proper decontamination of the affected area and equipment, and safe disposal of contaminated materials protect personnel and patients.

∴

## **DEFINITIONS.**

a. **Daily Cleaning.** Daily cleaning tasks are tasks conducted in suspect or confirmed EVD areas where the patient is receiving care. Patient-care staff will perform daily cleaning tasks to limit the number of people who enter patient-care areas.

b. **Terminal Cleaning.** Personnel conduct terminal cleaning after the MTF discharges the EVD patient from the area in order to disinfect the treatment area and medical equipment for safe care of new patients. Only designated cleaning personnel will perform terminal cleaning.

## **DESIGNATED PERSONNEL.**

The MTF Commander will determine whether in-house trained staff or contracted support personnel will conduct area disinfection. All designated cleaning personnel require training according to U.S. Army Medical Command (MEDCOM) and local infection control procedures on personal protective equipment (PPE) use and proper decontamination techniques.

## **PPE PROTOCOLS.**

Select and use PPE in all areas designated for terminal cleaning complying with hospital infection control standards and the Center for Disease Control and Prevention (CDC) guidance for EVD treatment. Train personnel to understand and mitigate the risks and hazards associated with the process. Ensure that personnel use proper PPE for selected cleaning and decontamination methods discussed in TERMINAL CLEANING AND DECONTAMINATION.

## **QUALITY ASSURANCE.**

Few references exist on how long Ebola viruses survive on surfaces. Viruses survive longer in liquid, within solids such as dried blood, in darkness, and in cold environments. For this reason, personnel implement more rigorous cleaning protocols in EVD patient-care areas. Clinical staff will clean the patient-care area, including high-touch surfaces like light switches, door knobs, and bed rails, constantly during patient treatment for their own protection.

a. **Cleaning Goals.** The goals of terminal EVD cleaning and disinfection are to: (1) dispose of any contaminated supplies and equipment; (2) dehydrate the room and remaining equipment; (3) manually clean and disinfect all applicable surfaces (directed by Infection Control) to remove all visible material; and

(4) disinfect all visibly clean surfaces and difficult to reach areas with ultraviolet light or encapsulation technology.

**b. Infection Control.** Infection Control personnel will assess and evaluate the decontamination procedures to ensure safety and effectiveness. Infection Control personnel will observe the decontamination process and provide oversight to ensure all items and areas are properly decontaminated and cleaned. Observation and oversight will include use of checklists, individual performance assessments, and necessary on-the-job corrections.

**c. Disinfectants.** Consult Infection Control to select an Environmental Protection Agency (EPA)-approved hospital disinfectant for use against the Ebola virus as recommended by the CDC<sup>1</sup> or make a 1:10 solution of bleach (1 part bleach added to 9 parts water) for terminal cleaning. Reference to a 1:10 bleach solution in this document assumes use of a 5 percent or greater chlorine strength bleach product.

## **DEVELOPING LOCAL PROCEDURES.**

Consider the following when developing local EVD terminal cleaning procedures:

**a. MEDCOM Guidance.** Seek clarification from Region and MEDCOM Logistics personnel about use of contract housekeepers for terminal cleaning tasks. The MEDCOM standing operating procedure, *Ebola Virus Disease Environmental Cleaning and Disinfection in the Medical Treatment Facility*, 31 October 2014, provides cleaning and disinfection techniques to conduct the terminal cleaning steps defined in TERMINAL CLEANING AND DECONTAMINATION.

**b. Proficiency in PPE Technique.** Clinical staff involved in EVD patient care may be best-suited for initial cleaning and disinfecting because they are the most familiar with the PPE procedures and the cleaning procedures implemented while the EVD patient was in the area.

**c. Contract Availability.** Housekeeping contract language may not allow cleaning medical equipment or isolation rooms. Purchase required technology or secure contract services in advance of need on a contingency basis.

---

<sup>1</sup><http://www.epa.gov/oppad001/list-l-ebola-virus.html>

**d. Personnel Willingness.** Personnel may refuse to enter isolation rooms because they are concerned about EVD exposure. Use effective risk communication, supply provisions, plans/procedures, and Command support to alleviate exposure concerns and improve morale of tasked individuals.

**e. Time.** Rapid turnaround of isolation rooms may be required to treat other patients once the EVD patient treatment ends. Develop a local protocol to permit deviation from the recommendations below (criteria for deviation, who has authority to issue, reviewers, revised cleaning goals, and so forth) when rapid turnaround of the treatment area is required.

**f. Compatibility of Medical Equipment and Facilities (Ventilation, Water, Safety) with decontamination materials and technologies.** Consult with Medical Maintenance, Industrial Hygiene, Safety, Facilities, Engineering, and other applicable personnel to evaluate potential damage to all medical equipment and facilities when subjected to disinfection chemicals and ultraviolet germicidal irradiation (UVGI).

**g. Room Layout and Ventilation Compatibility with Gas Disinfection and UVGI Treatment Technology.** Air must reach all room surfaces including interiors of cabinets and drawers and the undersides of the bed and equipment in the room for effective gas treatment. When equipment, furniture, and/or other items block the UV light, they reduce the effectiveness of UV disinfection. To increase UV disinfection effectiveness, perform UV disinfection multiple times until the UV light contacts all surfaces in the room.

#### **TERMINAL CLEANING AND DECONTAMINATION.**

Begin these steps after patient discharge from the room. For rooms used during initial entry into the MTF by patients under investigation for Ebola (such as emergency rooms, clinics, or transport vehicles), follow the steps a, c, d, e and i below for waste removal, floor cleaning, toilet cleaning, surface cleaning, and residual removal. For isolation rooms and areas used for extended patient care, follow all steps. Before cleaning, remove any sheeting and coverings that were put into place in a step-by-step process to avoid unnecessary cross contamination (i.e. ceiling first, next walls, then equipment, and so forth.). Begin terminal cleaning by removing all waste, blood, feces, dirt, etcetera because disinfection chemicals and UVGI are ineffective against EVD virus imbedded in organic matter.

**a. Remove Waste.** Remove and dispose as EVD waste all disposable equipment, supplies, linens, toiletries, food, and EVD waste. Terminal cleaners will generate EVD waste (PPE and cleaning supplies) during the remaining terminal cleaning process. Designated waste handlers will remain on hand in the cold zone to move EVD waste from the hot and warm zones to the EVD waste storage facility or the autoclave. Follow established EVD waste management procedures for the MTF (see Ebola Virus Disease Waste Management in the Medical Treatment Facility at:

<http://phc.amedd.army.mil/topics/discond/diseases/Pages/EbolaVirusDisease.aspx>).

**b. Initial Dehydration.** Adjust the ventilation system to achieve high air flows of 15-19 high-efficiency particulate air (HEPA)-filtered air exchanges per hour to desiccate any remaining viruses. Open cabinets and doors and otherwise provide airflow access. Seal the area in an undisturbed condition for 48 hours. After the 48-hour period, adjust ventilation to normal air flows and commence with manual cleaning procedures below. If rapid turnaround of the treatment area is required, enact the local protocol prescribed in DEVELOPING LOCAL PROCEDURES, paragraph e and document the decision.

**c. Surface Cleaning.** Clean all equipment and surfaces with bleach or approved disinfectant wipes. Clean all exterior and interior surfaces of equipment, cabinets, drawers, and fixtures (bathroom) with a 1:10 bleach solution or an EPA-approved hospital disinfectant for use against the Ebola virus as recommended by the CDC. Open cabinets and drawers, and clean the inside and outside surfaces. Treat all rooms and surfaces associated with EVD patient care regardless of the amount of visible contamination observed or the time passed since the patient left the area. Conduct the surface cleaning procedure twice. Collect all wipes and cleaning supplies as EVD waste.

**d. Toilet Cleaning.** Apply an EPA-approved hospital disinfectant for use against the Ebola virus or 1 cup of straight bleach solution around the bowl in the same manner as liquid/gel toilet bowl cleaner (i.e. apply to the inside top of the bowl and allow to run down into the bowl). Use 1 cup of disinfectant (or bleach). Allow 15 minutes of contact time, close the lid or cover, then flush toilet. After flushing, wipe the surfaces of the toilet (seat, handle, lid, inside bowl, outside of bowl, back, and so forth) with the approved disinfectant or a 1:10 bleach solution.

**e. Floor Cleaning.** Mop the floors twice with an EPA-approved hospital disinfectant for use against the Ebola virus as recommended by the CDC or a 1:10 bleach solution, starting with the cleanest areas (hallway) and ending with the most contaminated areas (patient room). Add an additional cup of the same approved hospital disinfectant or bleach solution used during the floor-cleaning

process, and allow 15 minutes of contact time. [Note: be careful not to mix incompatible disinfection chemicals during the cleaning process]. Pour treated mop water down the shower drain and pour an equivalent container size of water down the drain after the mop water. If a shower drain is not available, carefully pour into the toilet, close the lid or cover, and flush. Dispose of the mop heads, mop buckets, micro-fiber clothes, dusters, toilet brushes, and other cleaning devices as EVD waste. Dispose of mop and duster handles and frames as EVD waste. Mop handles and frames may be disinfected for reuse if: (1) personnel covered the mop handles and frames with plastic material prior to use, (2) personnel cleaned them with an EPA-approved hospital disinfectant for use against the Ebola virus as recommended by the CDC or a 1:10 bleach, and (3) Infection Control personnel approved them for reuse.

**f. Ultraviolet Germicidal Irradiation.** Acquire an UVGI-treatment device for use in the final stage of terminal cleaning, and follow the manufacturer's instructions for use and safety warnings. Mercury or xenon gas lamps generate ultraviolet C (UV-C) light. The UV-C light deactivates bacteria, viruses, and spores. The UV-C light may not extend into all the cracks and crevices in the EVD waste-storage area; however, it will reach all the surfaces touched by waste handlers. This treatment option is relatively quick, repetitious, and leaves no residue. Biosafety laboratories with built-in gas disinfection systems may activate the gas disinfection system instead.

(1) Ensure UV equipment operators use enough devices or conduct sufficient treatment cycles in the storage area to treat all exposed surfaces. Treat areas with a higher likelihood of contamination (entrance to the storage area) for a longer period to increase disinfectant exposure.

(2) Place treatment monitors under the equipment and at the darkest, furthest locations to ensure that the treatment reaches all areas.

(3) Assess the treatment area size to determine the appropriate amount of time to allow the by-products of the treatment process (ozone) to clear the area prior to re-entry. Activate the negative pressure system for the assessed time to remove any treatment bi-products.

**g. Terminal Cleaning Record.** Mark hand receipts of equipment dedicated to the EVD treatment room with the terminal cleaning date, time, and method. Affix a label or a tag to all portable equipment (such as pumps and pump stands) in the room and annotate the date and time of surface cleaning (step d) and UVGI/gas disinfection (step f).

**h. Repair Quarantine.** Hold disinfected equipment that requires repair in a two week quarantine status from the date of the UVGI treatment before initiating repairs.

**i. Clean to Remove Disinfectant Residue.** After completing all EVD terminal cleaning steps (paragraphs a–i), don standard hospital cleaning PPE to enter the area. Thoroughly clean the area and all equipment to remove chemical residuals left during the area disinfection process. Medical maintenance and other functional groups may retrieve decontaminated equipment from the room if necessary.

**j. Final Equipment Cleaning.** Medical maintenance and other functional groups will clean and disinfect all medical equipment removed from the quarantine according to manufacturer’s guidance and MTF policies by personnel wearing PPE. Select and manage cleaning supplies and PPE generated during this step according to routine housekeeping procedures.

#### **USE OF PROFESSIONAL SUPPORT.**

Numerous commercial companies specialize in hospital decontamination and offer the decontamination products/equipment and/or onsite decontamination support. If MTF personnel require outside decontamination products, equipment, and support, they should identify the procedures and products necessary and work with budget and contracting officials to coordinate acquisition and services in advance.

The table below identifies two professional organizations that identify certified local cleaning companies. State public health agencies may also regulate or recommend cleaning companies. Additional information obtained from MTFs and research organizations with experience in this subject are provided for reference. Entries in this table do not indicate an endorsement of a product or service by the U.S. Army.

<b>Name</b>	<b>Geographic Area and/or Treatment Method</b>	<b>Contact Information</b>
Institute of Inspection, Cleaning, and Restoration Certification (IICRC) Locate a Certified Professional listing	The IICRC is a certification and standard-setting non-profit organization for the inspection, cleaning and restoration industries. IICRC is developing an ANSI standard, BSR-IICRC S540 Crime Scene Cleanup.	<a href="http://www.iicrc.org/">http://www.iicrc.org/</a>
American BioRecovery Association listing	Nationwide, certification organization	888 979 2272, <a href="http://americanbiorecovery.org/">http://americanbiorecovery.org/</a> , or <a href="http://americanbiorecovery.org/SocialSite/MembersNearYou">http://americanbiorecovery.org/SocialSite/MembersNearYou</a>
National Institute of Health Biocontainment Unit	Hydrogen peroxide vapour (HPV) system from Bioquell	<a href="http://www.bioquell.com/en-uk/resources-and-support/product-support/sales-information-downloads/case-studies/case-study-all-categories/terminal-decontamination-biocontainment-unit-used-care-patient-ebola/">http://www.bioquell.com/en-uk/resources-and-support/product-support/sales-information-downloads/case-studies/case-study-all-categories/terminal-decontamination-biocontainment-unit-used-care-patient-ebola/</a>
University of Nebraska Medical Center Biocontainment Unit	UV germicidal irradiation (UVGI) generators from ClorDiSys Solutions, Inc. of Lebanon, New Jersey	<a href="http://www.clordisys.com/healthcareapp.php">http://www.clordisys.com/healthcareapp.php</a>
United States Army Medical Research Institute of Infectious Diseases (USAMRIID)	USAMRIID provided a general recommendation for decontamination of an EVD patient care area with a combined approach of Chlorine Dioxide (CD) and Ultraviolet (UV) light products and services by a company like ClorDiSys	<a href="http://www.clordisys.com/healthcareapp.php">http://www.clordisys.com/healthcareapp.php</a>
Madigan Army Medical Center	Xenex Pulsed Xenon UV Disinfection	<a href="http://www.xenex.com">www.xenex.com</a>

**STAFF MONITORING.**

Enforce proper monitoring of staff involved in terminal cleaning, handling, or management of contaminated waste, equipment, and materials. Maintain a list of the names, and contact information for all crews or individual workers involved in terminal cleaning duties for use by all members of the EVD Occupational Health or Safety team.

TIP No. 13-033-0616

**POINT OF CONTACT.**

The point of contact for this document is the Army Public Health Center (Provisional) Waste Management Program at 410-436-3651 or DSN 584-3651.

**Prepared by: Waste Management Program**

**Dated: June 2016**

Distribution A

## APPENDIX A

### REFERENCES

Army Public Health Center (Provisional). 2016. Environmental Health Engineering Portfolio, Standing Operating Procedure No. EHE37-001, Ebola Virus Disease Waste Management in the Medical Treatment Facility, 22 October 2014, revised June 2016, retrieved from <http://phc.amedd.army.mil/topics/discond/diseases/Pages/EbolaVirusDisease.aspx>

Bausch DG et al. 2007. Assessment of the Risk of Ebola Virus Transmission from Bodily Fluids and Fomites. *J of Infect Dis*;196:S142-S147.

Clay S, Maherchandani S, Malik YS, Goyal SM. 2006. Survival on uncommon fomites of feline calicivirus, a surrogate of noroviruses. *Am J Infect Control*. Feb;34(1):41-3.

EPA, Office of Pesticide Programs. 2009. January 9. List G, *EPA's registered antimicrobial products effective against norovirus (Norwalk-like virus)*. [http://www.epa.gov/oppad001/list\\_g\\_norovirus.pdf](http://www.epa.gov/oppad001/list_g_norovirus.pdf)

EPA, Office of Pesticide Programs, 2014. List L, *EPA's registered antimicrobial products that meet the CDC Criterial for use against the Ebola Virus*. <http://www.epa.gov/oppad001/list-l-ebola-virus.html>

Jelden KC, Gibbs SG, Smith PW, Schwedhelm MM, Iwen PC, Beam EL, Hayes AK, Marion N, Kratochvil CJ, Boulter KC, Hewlett AL, Lowe JJ. 2015. Nebraska Biocontainment Unit patient discharge and environmental decontamination after Ebola care. *American Journal of Infection Control*. Volume 43, Issue 3. March 2015. Epub 2015 January 27, 2015.

Lowe, John J, Gibbs, Shawn G, Iwen, Peter C, Smith, Philip W, and Hewlett, Angela L. 2013. Decontamination of a Hospital Room Using Gaseous Chlorine Dioxide: Bacillus anthracis, Francisella tularensis, and Yersinia pestis. *Journal of Occupational and Environmental Hygiene*, <http://dx.doi.org/10.1080/15459624.2013.818241>

Emory Healthcare, Draft Emory Healthcare Ebola Preparedness Protocols, October 20, 2014, retrieved from: [www.emoryhealthcare.org/ebolaprep](http://www.emoryhealthcare.org/ebolaprep)

TIP No. 13-033-0616

Lowe J, Gibbs S, Schwedhelm S, Nguyen J and Smith P. 2014. Nebraska Biocontainment Unit perspective on disposal of Ebola medical waste. *American Journal of Infection Control*. Volume 42, Issue 12.

Lowe JJ, Jelden KC, Schenarts PJ, Rupp LE, Hawes KJ, Tysor BM, Swansiger RG, Schwedhelm SS, Smith PW, Gibbs SG. 2014. Considerations for Safe EMS Transport of Patients Infected with Ebola Virus. *Prehospital Emergency Care: Official Journal of the National Association of EMS Physicians and the National Association of State EMS Directors*. November 7, 2014. PubMed Source: <http://www.ncbi.nlm.nih.gov/pubmed/25380073>

MEDCOM, G4 – Environmental Services, Standard Operating Procedure, *Ebola Virus Disease Environmental Cleaning and Disinfection in the Medical Treatment Facility*, 31 October 2014.

Piercy, TJ, Smither, SJ, Steward, JA, Eastaugh, L, Lever, MS. 2010. The survival of filoviruses in liquids, on solid substrates and in a dynamic aerosol. *Journal of Applied Microbiology*. 109(5):1531-1539, November.

Sagripanti JL, Lytle DC. 2011. Sensitivity to ultraviolet radiation of Lassa, vaccinia, and Ebola viruses dried on surfaces. *Arch Virol* 156:489-494

Sagripanti, JL, Rom, AM, Holland, LE. 2010. Persistence in darkness of virulent alphaviruses, Ebola virus, and Lassa virus deposited on solid surfaces. *Archives of Virology*. 155(12):2035-9, 2010 Dec.

U.S. Department of Health and Human Services. 2014. Centers for Disease Control and Prevention. Interim Guidance for Environmental Infection Control in Hospitals for Ebola Virus, October 2014. Retrieved on October 6, 2014 from <http://www.cdc.gov/vhf/ebola/hcp/environmental-infection-control-in-hospitals.html>

U.S. Department of Health and Human Services. 2014. Centers for Disease Control and Prevention. *Information for healthcare workers*. Retrieved from <http://www.cdc.gov/vhf/ebola/hcp/index.html>

U.S. Department of Health and Human Services. 2014. Centers for Disease Control and Prevention. *Protecting healthcare personnel*. Retrieved from <http://www.cdc.gov/HAI/prevent/ppe.html>

U.S. Department of Transportation. 2014. Pipeline and Hazardous Materials Safety Administration, DOT-SP 16279, October 17, 2014.

TIP No. 13-033-0616

World Health Organization. 2014. Interim infection prevention and control guidance for care of patients with suspected or confirmed filovirus hemorrhagic fever in health-care settings, with focus on Ebola.

<sup>1</sup><http://www.epa.gov/oppad001/list-l-ebola-virus.html>